

# Novel Scanometric Assay for Charantin in Bitter Melon (*Momordica charantia*) Extract Based on Immobilized Silver Nitrate and Methylene Blue as Colorimetric Paper

Muhammad Hilmi Afthoni<sup>1,2</sup>, Mariana Wahjudi<sup>3</sup> and Bambang Kuswandi<sup>2\*</sup>

<sup>1</sup>Department of Pharmacy, Faculty of Pharmacy, University of Surabaya, Raya Kalirungkut, Surabaya 60293 East Java, Indonesia

<sup>2</sup>Chemo and Biosensors Group, Faculty of Pharmacy, University of Jember, Jl. Kalimantan 37, Jember, 68121, Indonesia

<sup>3</sup>Faculty of Biotechnology, University of Surabaya, Raya Kalirungkut, Surabaya 60293 East Java, Indonesia

## \*Correspondence to:

Bambang Kuswandi  
Chemo and Biosensors Group  
Faculty of Pharmacy, University of Jember  
Jl. Kalimantan 37, Jember, 68121, Indonesia  
E-mail: [b\\_kuswandi.farmasi@unej.ac.id](mailto:b_kuswandi.farmasi@unej.ac.id)

**Received:** May 08, 2018

**Accepted:** July 03, 2018

**Published:** July 05, 2018

**Citation:** Afthoni MH, Wahjudi M, Kuswandi B. 2018. Novel Scanometric Assay for Charantin in Bitter Melon (*Momordica charantia*) Extract Based on Immobilized Silver Nitrate and Methylene Blue as Colorimetric Paper. *J Food Chem Nanotechnol* 4(2): 27-32.

**Copyright:** © 2018 Afthoni et al. This is an Open Access article distributed under the terms of the Creative Commons Attribution 4.0 International License (CC-BY) (<http://creativecommons.org/licenses/by/4.0/>) which permits commercial use, including reproduction, adaptation, and distribution of the article provided the original author and source are credited.

Published by United Scientific Group

## Abstract

Bitter melon (*Momordica charantia*) contains charantin that responsible for the antidiabetic activity. Therefore, it is important to determine charantin in the extracts for the purposed of its formulation or quality control. Herein, we proposed a novel scanometric assay for charantin determination in the bitter melon extracts based on immobilized silver nitrate and methylene blue as a colorimetric paper. The presence of charantin can be detected simply via the color change from blue to green and can be quantified using a scanner via the ImageJ software. The proposed charantin sensor has a response time of 8 mins with the linear range of 300 to 1000 µg/mL and the detection limit was 216.25 µg/mL. The sensor has good reproducibility (RSD = 2.86%) with the stability up-to 5 months. The sensor was applied to determine charantin in the sample extracts, and the result shown in a good agreement with TLC method.

## Keywords

Colorimetric sensor, Charantin, Silver nitrate, Methylene blue, Bitter melon

## Introduction

*Momordica charantia* or bitter melon commonly known as a fruit that is widely reported to have anticancer, antihyperlipidemic, analgesic, anti-inflammatory, antitumor, antiulcer, hepatoprotective, antioxidant activities [1], antidiabetic [2-8]. Charantin is a mixture of steroidal glycoside (1:1), and as an active compound or chemical marker contained in bitter melon that responsible for antidiabetic activity [4, 9]. Since charantin has the capability of lowered blood sugar equivalent to insulin [10]. Therefore, charantin has been used as an active compound or chemical marker for antidiabetic in the herbal medicine that has been formulated from the bitter melon extract [11]. Therefore, it is important to determine the charantin concentration in bitter melon extracts for the purposed of the antidiabetic formulation or quality control of charantin concentration as active compounds in the bitter melon extract used either as pharmaceutical or nutraceutical products purposes.

The conventional methods for determination of charantin are thin layer chromatography (TLC) [12], high performance thin layer chromatography (HPTLC) [13], high performance liquid chromatography (HPLC) [14], gases chromatography tandem mass spectroscopy (GC-MS) [15]. However, these methods are needed sample preparation, long procedure, and skilled personnel, as well as relatively expensive [16]. Certainly, for the big company or industry using conventional standard methods would not be an issue to comply with all

requirements needed, since they have enough resources, setting, and budget. However, for small and medium enterprises (SME), particularly home industry, this issue becomes a big problem for them, as they have a low resource, setting, and budget. Therefore, the development of the low-cost, portable and disposable method is needed for charantin detection in the bitter melon extracts particularly that can be used for SME including home industries. The aim of this study is to develop a simple method and low-cost for determination charantin as alternative compare to conventional standard methods. One alternative to the simple methods and low-cost is a chemical sensor. This is due to the fact that it can be developed as a colorimetric paper for simple strip or spot test by immobilization of suitable reagent solution for charantin onto a filter paper.

The reagents that can be used as an alternative that suitable for charantin detection is silver nitrate ( $\text{AgNO}_3$ ) coupled with methylene blue (MB). Since charantin could reduce  $\text{Ag}^+$  from  $\text{AgNO}_3$  into  $\text{Ag}^0$  as a silver nanoparticle (AgNPs), where AgNPs act as an electron transfer mediator and donates to MB as a redox catalyst termed as electron relay effect [17] and then reduce MB into leucomethylene blue (LMB) [18]. Based on this reaction, herein for the first time, we developed a novel scanometric assay for charantin in the bitter melon extract based on immobilized  $\text{AgNO}_3$  coupled with MB as colorimetric paper for simple detection of charantin in the bitter melon extracts. The colorimetric paper was fabricated simply by immobilizing the reagents mixture of  $\text{AgNO}_3$  and MB on the filter paper via adsorption method. The colorimetric paper results in the determination of charantin in the real sample of bitter melon extracts show a good agreement with the standard method using TLC. Thus, the results show that the proposed sensor can be used as a simple and rapid method for the determination of charantin in the bitter melon extracts.

## Materials and Methods

### Material

All reagents were used as purchased without further purification. Charantin was obtained from Xi'an Le Sen Biotechnology Co., Ltd (Shaanxi, China), methanol, ethanol chloroform, and hexane were purchased from Merck (Germany), silver nitrate ( $\text{AgNO}_3$ ), methylene blue (MB) and filter paper (Whatman cat No 1001 150) were purchased from Sigma Aldrich. Polyacrylic, nitrocellulose, aluminum foil obtained from Indomaret (Indonesia). Fresh *M. charantia* was purchased from Materia Medica (Malang, Indonesia). All chemicals used were of analytical reagent grade without further purification.

### Preparation of colorimetric paper

A colorimetric paper as charantin sensor was developed by immobilizing reagent solutions mixture of  $\text{AgNO}_3$  and MB onto a filter paper ( $\varnothing$  0,8 cm). The reagent immobilization was simply by using absorption method. This procedure was carried out by dipping the paper, that already cut into a circular shape (i.d. 5 mm), into 10 ml of the solution mixture overnight

(~12 hrs) at ambient temperature with a slow stirrer. Then, the immobilized reagent on the filter paper was dried at room temperature. Afterward, the sensor membrane was attached to white polyacrylic using double transparent tape (3M), to form a strip test for charantin. Then, the colorimetric paper was ready to be used as strip test for charantin detection as presented in figure 1A.




Figure 1: The colorimetric paper based on immobilized  $\text{AgNO}_3$  and methylene blue as a strip test is ready for use (A), and scanometric assay for charantin determination using the colorimetric paper as strip test (B).

### Sample preparation

The sample preparation was performed using previous method [19]. About 2 g sample was placed in conical flask extracted with 10 ml solvent consisting of methanol-water (80:20; %v/v). Then the sample was extracted using ultrasonic (Elmasonic, Germany) at 45 °C with ultrasonic power at 240 W for 2 hr. Afterword, the extract was filtered, then added with the methanol-water mixture until 10 ml. Then, the sample was ready to be used using the colorimetric paper.

### Optimization of colorimetric paper

The reagents concentration of  $\text{AgNO}_3$  and MB used in the colorimetric paper was according to previous method [18], from which the reagent concentrations were optimized. Here, the  $\text{AgNO}_3$  solution at concentration of 169 and 1690  $\mu\text{g}/\text{mL}$  were used, while MB at concentration of 32 and 1000  $\mu\text{g}/\text{mL}$  were used to make four combinations (1) 169:32; (2) 1690:32, (3) 169:1000, and (4) 1690:1000. The volume ratio (v/v) of selected reagents combination then optimized at the different volume ratio i.e. 8:2, 9:1 and 9.8:0.2 (v/v). Then, the sensor response time was also studied at the period from 0 to 20 min. For this study, 1000  $\mu\text{g}/\text{mL}$  of charantin was introduced into the sensor membrane. Lastly, the effect of the pH on the sensor response was also optimized between pH 5 to 8.

## Scanometric assay

For determination of charantin concentration, the scanometric assay was employed. Firstly, charantin solution was introduced into the sensor membrane, afterword waits at the desired time for the color change of the sensor to occur. Secondly, the color change of the sensors was captured by the flatbed scanner (CanoScan LIDE 110, Japan) at 300 dpi resolution. Finally, the scanned image was analyzed with ImageJ program (<https://imagej.nih.gov/ij/>). The intensity of the sensor color change is expressed as mean red, green and blue as well as RGB and presented as the intensity value of  $\Delta$  color unit of the sample subtracted with the blank in order to determine charantin concentration as given in figure 1B. All of the measurements were carried out in triplicate.

## Result and Discussion

### Sensing mechanism

The charantin detection mechanism is based on the reaction between charantin with  $\text{AgNO}_3$ , which in turn reduced MB. Herein, the reduction of MB is mediated by AgNPs formed as results of  $\text{AgNO}_3$  reduction by charantin in the bitter melon extract. This is according to Ajitha et al. [18], wherein this scheme, as a result of the decrease in MB absorbance, the increase of SPR peak of AgNPs was observed. In the electron transfer process, when there is a large redox potential difference between acceptor and donor, then there might be the restriction of electrons transfer. However, the electron transfer becomes easy, if an effective catalyst has intermediate redox potential value between acceptor and donor [20, 21]. Therefore, in this case, it reveals that AgNPs act as an electron transfer mediator since it accepts the electrons from extract and donates to MB by acting as a redox catalyst, which in turn termed as electron relay effect [17].

Similarly, in this sensing scheme, the charantin reduces  $\text{Ag}^+$  from  $\text{AgNO}_3$  into  $\text{Ag}^0$  as AgNPs, where AgNPs act as an electron transfer mediator and donates to MB as a redox catalyst termed as electron relay effect [17] and reduced MB into leucomethylene blue (LMB) [18] as presented in figure 2A. Based on this reaction scheme due to reduced MB into LMB, hence the sensor color change was from blue to green (Figure 2B). The color change was then analyzed using the ImageJ program for quantitative measurements, and the results show that the color value change or  $\Delta$  color unit in term of blue gave the best correlation coefficient ( $r$ ) compared to another color, i.e. red, green and RGB value (Figure 2C). Where  $\Delta$  color unit presents color of the sensor before and after reaction with charantin at various concentrations. In addition, the color change of the sensor in the  $\Delta$  blue was found to be the most proportional to the increasing concentration of charantin, followed by  $\Delta$ RGB. Therefore, the  $\Delta$  blue value was used for further measurements.

### Optimization of experimental parameters

Since the color change of the sensor response was from blue to green, when it tested with charantin (1000  $\mu\text{g/mL}$ ). In addition, the adsorption method of immobilization was




Figure 2: The sensing mechanism of the charantin sensor based on  $\text{AgNO}_3$  and methylene blue, where  $\text{Ag}^0$  (AgNPs) act as an electron transfer mediator, since it accepts the electrons from extract and donates to methylene blue by acting as a redox catalyst (A); the color change of the colorimetric sensor, the sensor before reaction; and after interaction with charantin (B); and the sensor response in the range of 300–1000  $\mu\text{g/mL}$  charantin, where the  $\Delta$ mean blue show the best correlation coefficient ( $r = 0.993$ ) (C). Here, the  $\Delta$  colour unit present color of the sensor before and after reaction with charantin at various concentrations. The experiments were carried out in triplicate.


employed in this case. Then, the optimizations of  $\text{AgNO}_3$  and MB concentrations for each of them in the mixture were needed to be performed as well as their volume ratio, as these value could be affecting the sensor response. In the case of concentration, the optimum concentration was found at 1690  $\mu\text{g/mL}$  and 1000  $\mu\text{g/mL}$  for  $\text{AgNO}_3$  and MB respectively as it gave optimum of the sensor response as given in figure 3A.

Furthermore, the optimization of volume ratio for  $\text{AgNO}_3$  and MB in the mixture was found at the volume ratio of 9.8:0.2 (v/v) for  $\text{AgNO}_3$  and MB respectively as this volume ratio gave optimum of the sensor response as shown in figure 3B.


Further experiment, the sample extract was used directly without any addition of buffer or other reagent and solution. In addition, this feature would add to the advantages of the proposed sensor, since make the sensor applications simpler and more practical particularly when it applied in the field or for small and home industries.

### Analytical characteristics

The response time of the sensor was observed within 20 minutes, where it gave a stable response toward charantin (Figure 4A). Here, all the time response selected (3, 5, 8 and 10 min) were give a good response as it gives correlation coefficient (R) close to 1. However, the response time of the sensor at 8 min was found to be optimum, where it gives the best linear sensor response towards charantin concentration as given in figure 4B. Based on this response time, the calibration curve was constructed and the sensor was found to have a linear concentration range in 300 to 1000  $\mu\text{g/mL}$  of charantin (Figure 4C), with coefficient correlation (r) was 0.993. The detection limit of the sensor response was



Beside the above parameters that would affect the charantin sensor prepared, another parameter that might affect the sensor response is pH. Here, the pH 5-7 was tested using pH buffer solution. In the case of pH, the charantin sensor response was not affected by the pH (5-7) as shown in figure 3C. Here, the pH 5 is similar to the pH of the charantin extract. Therefore, the sample tested no need to be conditioned with the pH buffer is usually used in many assays and simply can be used to determine the charantin concentration directly from the extract. This due to the fact that in the green synthesis of AgNPs, the pH buffer is not required [17, 18]. Thus, for



216.25  $\mu\text{g/mL}$  of charantin, and the quantification limit of the sensor was 648.76  $\mu\text{g/mL}$  of charantin. This limit of detection and the linear range are suitable to be used for the quality control of charantin in the bitter extracts, since the in-house herbal formulation was found to contain  $0.361 \pm 0.014$  of charantin (% w/w dry weight) or equal to 361  $\mu\text{g/mL}$ , while in the charantin content in *M. charantia* fruit extract was  $1.053 \pm 0.032$  (% w/w dry weight) or equal to 1053  $\mu\text{g/mL}$  [22]. In addition, the maximum yield of charantin on the dry weight basis of bitter melon was 3.18 mg/g or equal to 3180 mg/L. [1]. Thus, the proposed charantin sensor has a low limit of detection for quality control of charantin either in the herbal formulation or the dry bitter melon extracts.

In order to demonstrate the selectivity of the sensor, a selectivity study on the potential effect of other active compounds in the bitter melon extracts that may interfere the sensor response was studied, such as quercetin. The concentration ratio of 1:1 between charantin and quercetin were used (500  $\mu\text{g/mL}$ ). The interfering signal, calculated by subtracting the color intensity change of interfering substances added-sample with that of the original sample, and divided by color intensity change of original sample. The result shows that quercetin interferes charantin with the interference value of 43.769%. This is due to the fact that quercetin also reduces  $\text{AgNO}_3$  to become AgNPs [23, 24]. Therefore, the sample preparation has to be separate quercetin from the extract. Here, the quercetin could be separate from the extract by degradation of quercetin via ultrasonication of the extract, so that its reduction activity will be reduced [25]. Thus, the sample could be prepared by ultrasonication during 2 hrs at 45  $^\circ\text{C}$ , and then extracted with the methanol-water mixture (80:20% v/v). After this degradation step, the quercetin interference was less than 5%, and the other hand interfering signal from pharmaceutical excipient used sucrose, nipagin, starch respectively as show in figure 5. The interference of pharmaceutical excipient less than 5% and this value was acceptable for this type of assay [26].




Figure 5: The possible interferences arising from pharmaceutical excipients.

In term of the reproducibility using six sample at the concentration 700  $\mu\text{g/mL}$ , the sensor shows good reproducibility with the RSD value was calculated to be 2.868%. This reproducibility value was indicated that the

sensor has the good precision for the charantin determination [26]. The accuracy of the sensor was performed using standard addition method, where the recovery values (%) for the accuracy of the charantin sensor was calculated to be 100.605%. Thus, this value complies with the recovery values within 0.1% of sample concentration tested (95-105%) [26]. Therefore, this value indicated that this method has a good accuracy in determination of charantin concentration.

### Stability

The stability study of the sensor response was carried out at the room temperature (25  $^\circ\text{C}$ ). The sensor response was evaluated every day until 10% of the sensor response decreased from the initial response value. The result shows that the sensor was very stable at room temperature (25  $^\circ\text{C}$ ) up to 5 months, since afterword, the sensor response was observed to be decreased more than 10% at room temperature.

### Application of the real samples

In order to demonstrate the applicability of the proposed sensor in the determination of charantin in the real bitter melon extracts. The sensor was applied for the detection of charantin in the real sample extracts. The content of charantin in the bitter melon extracts was calculated based on the calibration curve constructed above. The results of the proposed charantin sensor were also compared with the result determined using the standard method, i.e. TLC – densitometry method [27] and the results are summarized in table 1. Based on these two results are found to be no difference between the two methods, since the calculated p-value were more than  $\alpha = 0.05$ . Thus, it can be stated that the proposed charantin sensor is in good agreement with the standard method (TLC), and can be suggested that the proposed charantin sensor can be used as an alternative method for the charantin determination in the bitter melon extract samples.

Table 1: The results of charantin determination in the bitter melon extract samples using the proposed charantin sensor and TLC-densitometric method (n = 3,  $\alpha = 0.05$ ).

Sample	Charantin Sensor	TLC - Densitometer	P(T ≤ t) two-tail
A	0.286 ± 0.010	0.288 ± 0.0035	0.723
B	0.241 ± 0.006	0.232 ± 0.0035	0.135
C	0.313 ± 0.010	0.314 ± 0.0007	0.869

### Conclusion

The novel scanometric assay based on a colorimetric paper of immobilized reagent mixture of  $\text{AgNO}_3$  and MB has been developed for measuring charantin in the bitter melon extract. Here, the charantin sensor has a linear response at 300-700  $\mu\text{g/mL}$  with the limit of detection at 216.25  $\mu\text{g/mL}$  and the results were found to be reproducible. The proposed method is simple, easy to operate, and low-cost, since it is fabricated as a colorimetric paper in the form of the strip test, making it as a simple and rapid method for determination charantin in the bitter melon extracts.

## Acknowledgments

The authors thank the Dr. Soebandi Vocational School, Jember, for supporting this work via study leave to MHA, and Lediana Sari and Venny Dyah for helping in providing the bitter melon samples.

## References

1. Ahamad J, Amin S, Mir SR. 2015. Simultaneous quantification of gymnemic acid as gymnemagenin and charantin as  $\beta$ -sitosterol using validated HPTLC densitometric method. *J Chromatogr Sci* 53(7): 1203-1209. <https://doi.org/10.1093/chromsci/bmu166>
2. Budrat P, Shotipruk A. 2008. Extraction of phenolic compounds from fruits of bitter melon (*Momordica charantia*) with subcritical water extraction and antioxidant activities of these extracts. *Chiang Mai Journal of Science* 35(1): 123-130.
3. Aswar PB, Kuchekar BS. 2012. Phytochemical, microscopic, antidiabetic, biochemical and histopathological evaluation of *Momordica charantia* fruits. *Int J Pharm Pharm Sci* 4(1): 325-331.
4. Upadhyay A, Agrahari P, Singh DK. 2015. A review on salient pharmacological features of *Momordica charantia*. *International Journal of Pharmacology* 11(5): 405-413. <https://doi.org/10.3923/ijp.2015.405.413>
5. Joseph B, Jini, D. 2013. Antidiabetic effects of *Momordica charantia* (bitter melon) and its medicinal potency. *Asian Pac J Trop Dis* 3(2): 93-102. [https://doi.org/10.1016/S2222-1808\(13\)60052-3](https://doi.org/10.1016/S2222-1808(13)60052-3)
6. Tahira S, Hussain F. 2014. Antidiabetic evaluation of *Momordica charantia* L. fruit extracts. *West Indian Med J* 63(4): 294-299. <https://doi.org/10.7727/wimj.2013.180>
7. Perumal V, Khoo WC, Abdul-Hamid A, Ismail A, Saari K, et al. 2015. Evaluation of antidiabetic properties of *Momordica charantia* in streptozotocin induced diabetic rats using metabolomics approach. *Int Food Res J* 22(3): 1298-1306.
8. Ahmed I, Lakhani MS, Gillett M, John A, Raza H. 2001. Hypotriglyceridemic and hypocholesterolemic effects of anti-diabetic *Momordica charantia* (karela) fruit extract in streptozotocin-induced diabetic rats. *Diabetes Res Clin Pract* 51(3): 155-161. [https://doi.org/10.1016/S0168-8227\(00\)00224-2](https://doi.org/10.1016/S0168-8227(00)00224-2)
9. Hazarika R, Parida P, Neog B, Singh Yadav RN. 2012. Binding energy calculation of GSK-3 protein of human against some anti-diabetic compounds of *Momordica charantia* linn (Bitter melon). *Bioinformation* 8(6): 251-254. <https://doi.org/10.6026/97320630008251>
10. Desai S, Tatke P. 2015. Charantin: an important lead compound from *Momordica charantia* for the treatment of diabetes. *J Pharmacogn Phytochem* 3(6): 163-166.
11. Gautam A, Kashyap SJ, Sharma PK, Garg VK, Visht S, et al. 2010. Identification, evaluation and standardization of herbal drugs: a review. *Scholars Research Library* 2(6): 302-315.
12. Saifi A, Namdeo KP, Chauhan R, Dwivedi J. 2014. Evaluation of pharmacognostical, phytochemical and antidiabetic activity fruits of *Momordica charantia* Linn. *Asian J Pharm Clin Res* 7(3): 152-156.
13. Shanmugapriya R, Poornima S. 2014. Detection of charantin in the leaves and fruits of *Momordica tuberosa* (Cogn) Roxb and *Momordica dioica* (Roxb Ex Wild) by Analytical HPTLC. *International Journal of Scientific and Research Publications* 4(6): 1-8.
14. Alam S, Asad M, Asdaq SM, Prasad VS. 2009. Antiulcer activity of methanolic extract of *Momordica charantia* L. in rats. *J Ethnopharmacol* 123(3): 464-469. <https://doi.org/10.1016/j.jep.2009.03.024>
15. Singh R, Kumar A, Giri DD, Pandey KD. 2012. Gas chromatography-mass spectrometry analysis and phytochemical screening of methanolic fruit extract of *Momordica charantia*. *J Rec Adv Agri* 1(4): 122-127.
16. Reis PB, Ramos RM, de Souza LF, de Vasconcelos Cançado S. 2009. Validation of spectrophotometric method to detect and quantify nitrite in ham pate. *Braz J Pharm* 45(1): 49-55. <https://doi.org/10.1590/S1984-82502009000100006>
17. Gupta N, Singh HP, Sharma RK. 2011. Metal nanoparticles with high catalytic activity in degradation of methyl orange: an electron relay effect. *J Mol Catal A Chem* 355(1-2): 248-252. <https://doi.org/10.1016/j.molcata.2010.12.001>
18. Ajitha B, Reddy YAK, Reddy PS. 2015. Biosynthesis of silver nanoparticles using *Momordica charantia* leaf broth: evaluation of their innate antimicrobial and catalytic activities. *J Photochem Photobiol B* 146: 1-9. <https://doi.org/10.1016/j.jphotobiol.2015.02.017>
19. Ahamad J, Saima A, Mir SR. 2015. Optimization of ultrasound-assisted extraction of charantin from *Momordica charantia* fruits using response surface methodology. *J Pharm Bioallied Sci* 7(4): 304-307. <https://doi.org/10.4103/0975-7406.168032>
20. Mallick K, Witcomb M, Scurrell M. 2006. Silver nanoparticle catalysed redox reaction: an electron relay effect. *Mater Chem Phys* 97(2-3): 283-287. <https://doi.org/10.1016/j.matchemphys.2005.08.011>
21. Kundu S, Ghosh SK, Mandal M. 2002. Silver and gold nanocluster catalyzed reduction of methylene blue by arsine in micellar medium. *Bull Mater Sci* 25(6): 577-579.
22. Ahamad J, Amin S, Mir SR. 2014. Development and validation of HPTLC densitometric method for estimation of charantin in *Momordica charantia* fruits and herbal formulation. *J Pharmacogn Phytochem* 2(5): 126-130.
23. Vanaraj S, Keerthana BB, Preethi K. 2017. Biosynthesis, characterization of silver nanoparticles using quercetin from *Clitoria ternatea* L to enhance toxicity against bacterial biofilm. *J Inorg Organomet Polym Mater* 27(5): 1412-1422. <https://doi.org/10.1007/s1090>
24. Jain S, Mehata MS. 2017. Medicinal plant leaf extract and pure flavonoid mediated green synthesis of silver nanoparticles and their enhanced antibacterial property. *Scientific Reports* 7(1): 1-13. <https://doi.org/10.1038/s41598-017-15724-8>
25. Hemwimon S, Pavasant P, Shotipru A. 2007. Microwave-assisted extraction of antioxidative anthraquinones from roots of *Morinda citrifolia*. *Sep Purif Technol* 54(1): 44-50. <https://doi.org/10.1016/j.seppur.2006.08.014>
26. Huber L. 2007. Validation and qualification in analytical laboratories. 2<sup>nd</sup> edition. Informa Healthcare, New York, USA.
27. Thomas CT, Reddy YP, Devanna N. 2012. Impact of cooking on charantin estimated from bitter melon fruits using high performance thin layer chromatography. *International Research Journal of Pharmacy* 3(6): 149-154.